

**CORRIGENDUM FOR TENDER LISTED UNDER SR. 4 OF SHORT  
TERM GLOBAL TENDER NOTICE NO: 16/2015**

**Tender No: 14-VI/PKS(788)2015PB/T-120 Procurement of UHV Magnetron Sputtering system- correction in specification-reg.**

Due to typographical mistake in technical specifications following corrigendum is issued as under. Other terms, conditions and specifications of said tender will remain as the same.

Tender NO:	Existing Specification of tender under Sr. 4.2	<b>Correct Specification is to be read as under Sr. 4.2</b>
14-VI/PKS(788)2015PB/T-120	Turbo molecular pump for load lock along with compatible Rotary pump to achieve <b><u>1x10<sup>-8</sup>mbar</u></b> vacuum in max. 15 minutes.	Turbo molecular pump for load lock along with compatible Rotary pump to achieve <b><u>1x10<sup>-6</sup> mbar</u></b> vacuum in max. 15 minutes.

**Last date of receipt of complete tender & date of opening of technical bid for Tender No: 14-VI/PKS(788)2015PB/T-120 only may be read as 18.03.2016 upto 1.00PM (IST) & Opening of tender will be 18.03.2016 at 2.30 PM (IST) onwards.**

**Controller of Stores & Purchase**